


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-S-P-EA-EW*

S.N.: *7552*

Date: *MAR 18 1996*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: ($R_L = 50\Omega$)
50 TO 500 NS
- c) Rise time: *0.5% MAX DUTY CYCLE*
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 100 KHz*
0.5% MAX DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power:

+24 VDC
600 mA MAX

TRF $R_L = 50\Omega$, *50 V/DIV*
100 NS/DIV

BOI *6066 ATTN*
100 VOLTS/DIV
10 NS/DIV.

[Signature]