


PULSE GENERATOR
PERFORMANCE CHECK

Model: ^{HVM} ~~AVM~~ -2-C-POT-M

S.N.: 7556

Date: DEC 12 1995



- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
 ≤ 100 PS
- d) Fall time:
 ≤ 135 PS
- e) PRF: 0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V,
50-60 Hz.
RF

40 dB ATTEN

10 VOLTS DIV

PRF \approx 12 MHz