


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-441-C-N*

S.N.: *7576*

Date: *MAR 11 1996*



- a) Output signal amplitude: *0 TO -50 VOLTS*
- b) Pulse width: *50 NS*
2 TO 100 NS
- c) Rise time: *≤ 250 PS*
- d) Fall time: *≤ 2 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100/200 VOLTS*
(± 10%)
50-60 Hz
✓

TRP 10 NS/DIV

PROT 500 PS/DIV
(RISE TIME)

50 dB ATTN

2 32 VOLTS/DIV