


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-N*
S.N.: *7579*
Date: *FEB 6 1986*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0.2 TO 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

+24 VDC
600 mA (MAX)
g

40 db ATTEN

10 VOLT/DIV

PRF = 10 MHz.