


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PA-FD1-PS-M1*

S.N.: *7589*

Date: *MAR 13 1996*



- a) Output signal amplitude: *+15 VOLT, R_L 50Ω*
- b) Pulse width: *50 ns to 50 μs*
- c) DIVISION: *Factor: 2 to 999*
- e) PRF: *0 to 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100/200 VOLT
± 10%
50-60 Hz*

TDP input ≈ 10 MHz⁺

*BOI out, $\frac{1}{10}$
R_L = 50 Ω*

