


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMH-121-PSV-55C1*

S.N.: *7594*

Date: *JAN 29 1996*



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA MAX
- h) t_p, t_n : *TEMPERATURE*
MATCHED

60 db ATTEN
100 VOLTS/DIV
PRF ≈ 30 KHz

MS