


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-121-PN-554*

S.N.: *7595*

Date: *JAN 29 1996*



60 db ATTEN

100 VOLTS/DIV

PRF ≈ 30 KHZ

- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability:
OK
- g) Prime power: *+15 VDC*
250 mA MAX
- h) *t_p, t_n: TEMPERATURES*
MATCHED.

[Signature]