


PULSE GENERATOR
PERFORMANCE CHECK

AVH-HUI-DW-SSC

: 7596

ate: DW 30 1996



a) Output signal amplitude:

$\pm 100 \text{ V}$

b) Pulse width:

0.6 TO 2.0 NS

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 KHz

f) Jitter, stability:

OK

g) Prime power: $\pm 15 \text{ VDC}$

250 mA MAX

h) t_r, t_f : TEMPERATURE
UNSTATED

60 dB ATTN.

100 VOLTS/DIV

PRF $\approx 30 \text{ KHz}$