


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-421-DN-55C1*

S.N.: *7598*

Date: *JAN 31 1996*



- a) Output signal amplitude:
± 100 VOLT
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power: *+15 VDC*
250 mA MAX
- h) *t_p, t_r*: *TEMPERATURE*
AUTOMATIC

60 dB ATTEN

100 VOLT/DIV

PRF = 30 KHz

[Signature]