


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7603

Date: FEB 12 1996



- a) Output signal amplitude:
 ± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS

250 mA MAX

TOP: P OUT

BOT: N OUT

60 dB ATTEN

100 VOLTS/DIV

PRF = 30 kHz