


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-3-P-M1*

S.N.: *7610*

Date: *FEB 16 1996*



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
650 TO 700 ps
- c) Rise time: *(FWHM)*
NA
- d) Fall time:
NA
- e) PRF: *0 TO 20 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

+ 24 VDC
600 mA MAX

50 dB ATTEN

= 32 VOLTS / DIV

PRF = 10 MHz.