


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7622

Date: APRIL 1 1996



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
250 mA MAX

TOP: P OUT
BOT: N OUT
60 dB ATTEN
100 VOLTS/DIV
PRF = 30 KHZ