


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-3-PW-C-P-BTIA*

S.N.: *7625*

Date: *MAY 27 1996*



- a) Output signal amplitude:
0 TO +100 V
- b) Pulse width: *($R_L > 50 \Omega$)*
0.5 TO 50 ms
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF: *0.2 TO 2.0 Hz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

$R_L = 50 \Omega$
PRF = 2 Hz

120/240 V
50-60 Hz

[Signature]