


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MV0-9D-C-P*

S.N.: *7628*

Date: *MAR 27 1976*



- a) Output signal amplitude:
0 TO +100 MA
- b) Pulse width:
10 TO 200 NS
(30% max duty cycle)
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 1 NS
- e) PRF: *0 TO 25 MHz.*
- f) Jitter, stability:
O.K.
- g) Prime power:
120 / 240 V,
50 - 60 kHz

MAIN FREQUENCY OUT TO
50 ~ (40 db)

TOP *5 NS/DIV, PRF = 3 MHz*

BOT *10 NS/DIV, PRF = 25 MHz*