


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M-1002-C*

S.N.: *7632*

Date: *Feb 13 1991*



- a) Output signal amplitude:
*0 TO ± 10 VOLTS
(+ TTL + ECL)*
- b) Pulse width:
50 NS TO 0.5 SEC
- c) Rise time:
6.5 NS
- d) Fall time:
6.5 NS
- e) PRF:
1 Hz TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

*TOP 50 ns/DIV, 5V/DIV
PRF ≈ 10 MHz
DC OFFSET "OFF"*

*POT 1 ms/DIV, 5V/DIV
-5V PULSE WITH
+2.5V DC OFFSET
R_L = 50 Ω*

[Signature]