


PULSE GENERATOR
PERFORMANCE CHECK



Model: *AVO-6C1-C-P-M-0P1*

S.N.: *7636*

Date: *FEB 28 1991*



- a) Output signal amplitude: *0 TO +5 AMP*
- b) Pulse width: *50NS TO 50US*
- c) Rise time: *(1% OUT CYCLE MAX) ≤ 10 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 10 KHZ (1% MAX OUT CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 HZ



(B) AS (A) BUT 10US/DIV
+ PRF ≈ 200 HZ

[Signature]