


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MV-14V3-C-CA-EW-M-T*
S.N.: *7640*
Date: *MM 8 1992*



- a) Output signal amplitude:
0 TO 40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 - 60 Hz

50 dB ATTEN

32 VOLTS/DIV

PRF = 100 kHz

✓