


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-1-e-p*

S.N.: *7643*

Date: *MAY 23 1996*



- a) Output signal amplitude:  
*0 TO +50 VOLTS*
- b) Pulse width:  
*0.2 TO 1.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLT  
50-60 HZ*

*40 db ATTEN  
25 VOLTS/DIV.*

*UNIT SHIPPED  
WITH VOLTAGE  
SELECTOR CHANGED  
IN THE 240V  
POSITION.*

*[Signature]*