


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVRH-3-C-PN*

S.N.: *7644*


Date: *JUNE 24 1986*



- a) Output signal amplitude:
0 TO ± 3000 VOLTS
- b) Pulse width: *(R_L > 10Ω)*
0.25 TO 2.545.
- c) Rise time:
≤ 100ns
- d) Fall time:
≤ 100ns
- e) PRF: *0 TO 1 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:

P_{out} ≈ 500V/DIV.
PRF = 1KHz

120/240 VOLTS
50-60Hz.



[Signature]