


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-4A-P*

S.N.: *7647*

Date: *JUNE 20 1996*



- a) Output signal amplitude:  
*0 TO +650 VOLTS*
- b) Pulse width: *( $t_r = 50\text{ns}$ )*  
 *$\leq 5\text{NS}$  (20% point)*
- c) Rise time:  
 *$\leq 1\text{NS}$*
- d) Fall time:  
 *$\leq 3\text{NS}$*
- e) PRF:  
*0 TO 10 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VDC*

*800 mA MAX.*

*60 db ATTEN*

*100 VOLTS/DIV*

*2 NS/DIV*

*PRF = 10 KHz.*

*[Signature]*