


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MU-141C-PS*


S.N.: *7654*

Date: *MAY 9 1986*



- a) Output signal amplitude: *1 VOLT MAX.*
- b) GAIN: *> 20 db.*
- c) Rise time: $\leq 1 \text{ NS}$
- d) Fall time: $\leq 1 \text{ NS}$
- e) PRF: *0 TO 600 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/200V*
50-60W

(A) OUTPUT TO 20 db
TOP 10 NS/DIV
BUT 1 NS/DIV (RISE TIME)



OUTPUT TO 50W
100 NS/DIV
WIDE PULSES TEST.

(B)