


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P.*
S.N.: *7658*
Date: *APRIL 17 1996*



- a) Output signal amplitude:
+25 TO +100 mV
- b) Pulse width:
≤ 25 ns
- c) Rise time: *(HARF Power)*
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK

- g) Prime power:
+15 VDC
185 mA (max)
- [Signature]*

≈ 60 mV/div
PRF = 1.0 kHz