


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-6A-PS-N-PTA*


S.N.: *7659*

Date: *MAY 28 1996*



- a) Output signal amplitude:
0 TO -1.0 AMP
- b) Pulse width:
50 TO 500 NS
- c) Rise time:
≤ 5 NS
- d) Fall time:
≤ 5 NS
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

① *TOP: INPUT TRIGGER TO MAIN FRAME (5V/div)*
BUT: OUTPUT TO 50Ω (10V/div)



② *AVX-S1 ME OUT TO 50Ω (1V/div)*