


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-108A-2-N-C-072*
 S.N.: *7663*
 Date: *JUN 23 1996*



- a) Output signal amplitude: *0 TO -100 AMPS*
(0 TO -50 V)
- b) Pulse width: *100US TO 1 MS*
- c) Rise time: *(0.8% MAX DUTY CYCLE)*
≤ 5US
- d) Fall time: *≤ 5US*
- e) PRF: *0 TO 1 KHZ*
- f) Jitter, stability: *(0.8% MAX DUTY CYCLE)*
OK
- g) Prime power: *120/240 V*
50-60 HZ

TOP MONITOR OUT
TO LK. (100 AMP)

POT. LOAD VOLTAGE,
 $R_L = 0.1 \Omega$
100 AMP)

HORZ; 20US/DIV

