

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVRH-3-C-PN-AT-PWT-MHV

S.N.: 7685

Date: JULY 11 1986

a) Output signal amplitude:
0 TO ± 3 KV, $R_L \geq 10K$.

b) Pulse width:
200 NS \rightarrow 2.5 μ S


c) Rise time:
 ≤ 100 NS

d) Fall time:
 ≤ 100 NS

e) PRF: 0 TO 1 KHz.

f) Jitter, stability:
OK

g) Prime power:
100 VOLTS,
50-60 Hz



≈ 500 VOLTS / DIV
500 NS / DIV
 $R_L \approx 10K$.

