


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NH-5A-C-P*

S.N.: *7686*

Date: *JULY 26 1996*



- a) Output signal amplitude: *0 to +20V*
- b) Pulse width: *200 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

40 db Attenu
10 volts/div
TOP 1 ns/div
BOT 200 ps/div
PRF = 1.0 MHz

50-60 MHz
[Signature]