


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *M-144.B1*

S.N.: *7692*

Date: *JULY 25 1996*



- a) Output signal amplitude:  
*+ 20 VOLTS TO 50V*
- b) Pulse width:  
*10NS TO DC*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 2 NS*
- e) *INPUT TRIGGER: TTL*
- f) Jitter, stability: *OK*
- g) Prime power:  
*+ 28 ± 1.0 VOLTS*  
*600 mA (MAX)*

*40 db ATTN*

*10 VOLTS / DIV.*

*PRF = 10 KHz.*