


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-L-P*

S.N.: *7694*

Date: *SEPT 23 1996*



- a) Output signal amplitude: *0 to +100 VOLTS*
- b) Pulse width: *≤ 1.0 ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

± 60db ATTEN

100 VOLTS/DIV

PRF = 1.0 MHz

[Signature]