


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MA-MVI-C-T1*

S.N.: *7722*

Date: *JULY 22 1996*



- a) Output signal amplitude: *0 TO +100V*
- b) Pulse width: *(TO 50ns)*  
*≤ 600PS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60Hz*

*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 100 kHz*

*[Signature]*