


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM4-1-P*

S.N.: *7729*

Date: *June 10 1996*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 0.2 NS
- c) Rise time:
≤ 100PS
- d) Fall time:
≤ 100PS
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

+24 VDC
500 mA (max)
[Signature]

40 db ATTEN
10 VOLTS/DIV
TOP 1 NS/DIV
BOT 200 PS/DIV
210 MHz PRF