




PULSE GENERATOR
PERFORMANCE CHECK

Model: *M-1011-C*
 S.N.: *7734*
 Date: *SEPT 10 1996*



- a) Output signal amplitude: *0 TO ± 100 VOLTS*
- b) Pulse width: *($R_L \geq 50 \Omega$)
100 NS TO 1.0 MS
(10% MAX DUTY CYCLE)*
- c) Rise time: *≤ 10 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 1.0 MHz
(10% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/200V, 50-60 Hz*



*(B) Nout: $R_L = 50 \Omega$ 10V/DIV
 $R_L = 50 \Omega$ 500 NS/DIV*