


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ADA-FD2-PS-IP*

S.N.: *7735*

Date: *SEPT 20 1996*



a) Output signal amplitude: *772 TO 50mV*

b) Pulse width: *10NS TO 5us*

c) Rise time: *NA*

d) Fall time: *NA*

e) PRF: *100 Hz TO 100 MHz*

f) Jitter, stability: *OK*

g) Prime power: *120/240V*

h) DIVISION FACTOR: *50-68 Hz*  
*2 TO 255*

*DIVISION FACTOR = 100*

*[Signature]*