


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1-P

S.N.: 7737

Date: JUNE 19 1996



- a) Output signal amplitude:
+25 TO +100 Amp
- b) Pulse width:
≤ 25 ns (HALF power)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK

≈ 60 Amp / DIV
PRF = 1.0 KHz

- g) Prime power:
+15 VDC
165 mA (MAX)
NA