


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-F*

S.N.: *7738*

Date: *JUNE 19 1996*



- a) Output signal amplitude:
+ 25 V + 100 Amp;
- b) Pulse width:
≤ 25 ns (1/2 power)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 kHz
- f) Jitter, stability:
OK.
- g) Prime power:
+ 15 VDC
10.5 mA (max)

≈ 60 mV/div
PRF ≈ 1.0 kHz

[Signature]