


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-155C-C-P-M-ET-OT-PWT*


S.N.: *7739*

Date: *SEPT 30 1996*




- a) Output signal amplitude:
PULSE: 0 TO +2.0 AMP
- b) DC: *0 TO 50 mA*
Pulse width:
100 NS TO 10 US
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK

① *1V/DIV, 500 NS/DIV*



- g) Prime power: *120 / 240 V*
50 - 60 Hz

② *1V* ③ *BUT 2US/DIV*



*DC OFFSET
SET AT 500 mA*