


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-5A1-C-N*

S.N.: *7771*

Date: *OCT 24 1996*



- a) Output signal amplitude:
0 TO 10 AMP
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 10 AMP/Div

R_L ≈ 10 OHMS.

PRF ≈ 1 KHz

120/240 V
50-60 Hz.
AK