


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ML-2A C-W-P*

S.N.: *7781*

Date: *OCT 10 1996*



- a) Output signal amplitude: *0 to +170 V*
- b) Pulse width: *($R_L = 50 \Omega$)
3 to 400 ns**
- c) Rise time: *≤ 2 ns*
- d) Fall time: *6.3 ns*
- e) PRF: *0 to 5 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 Hz*

60 db ATTEN
100 VOLT/DIV
PRF = 5.0 kHz

** FOR 100-400mV
REQUIRES 100'
OR 50 Ω
CABLE BETWEEN
A & B TERMINALS.*