


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-HV3-P-55A*

S.N.: *7786*

Date: *OCT 21 1996*



- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width:  
*0.6 TO 4.0 NS*
- c) Rise time:  
*≤ 150 PS*
- d) Fall time:  
*≤ 600 PS*
- e) PRF: *0 TO 500 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *+15 VDC*  
*250 mA (MAX)*

*50 dB ATTEN*  
*1A 32 VOLTS/DIV*  
*PRF = 300 KHz*

*[Signature]*