


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR4-1-PS-N-A52

S.N.: 7801

Date: DEC 20 1996



- a) Output signal amplitude: 0 TO -1000 V
- b) Pulse width: ($R_L \geq 1000\Omega$)
0.1 TO 10 μ S
- c) Rise time: ≤ 50 ns
- d) Fall time: ≤ 50 ns
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability: OK
- g) Prime power: 170/240 V
50-60 Hz

TOP INPUT TRIGGER
5 V/DIV, 2 μ S/DIV

BOT OUTPUT
 ≈ 500 V/DIV, 2 μ S/DIV
 $R_L = 10$ K

[Signature]