


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APRH-3-PS-PN-PWT*
 S.N.: *7813*
 Date: *NOV 12 1996*



- a) Output signal amplitude:
0 TO ± 3000 V
(R_L > 10K)
- b) Pulse width:
0.2 TO 2.5 μs
- c) Rise time:
≤ 100 NS
- d) Fall time:
≤ 100 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz

500 NS/Div
= 500 V/Div

