


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMMP-2CP*
S.N.: *7823*
Date: *NOV 11 1996*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTN
10 VOLTS/DIV
TOP: 5 NS/DIV
BOT: 200 PS/DIV
(RISE TIME)

120 / 240 V
50 - 60 Hz
AF