

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-C-BE03-R5-P-0P1-M*

S.N.: *7845*

Date: *NOV 18 1996*

a) Output signal amplitude:

*0 TO +2A*

b) Pulse width:

*17 NS*

c) Rise time:

*≤ 2 NS*

d) Fall time:

*≤ 2 NS*


e) PRF:

*0 TO 20 KHz*

f) Jitter, stability:

*OK*

g) Prime power: *120/240V*  
*50-60Hz*



*MONITOR OUT TO*  
*40 Ohms (100 V/DIV)*  
*2 AMP/DIV*

*AVO-2T-BE02B3 CONNECTED*  
*VIA 5 METER CABLE*  
*IN 45.9A IN SERIES WITH*  
*1 OHM 4500 OHM TEST LOAD*  
*PRF = 20 KHz*