


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-M-1-PS-POS*

S.N.: *7849*

Date: *JAN 06 1996*



- a) Output signal amplitude:
0 to +100 V_{0.93}
- b) Pulse width:
3 to 100 ns
- c) Rise time:
≤ 1 ns
- d) Fall time:
≤ 2 ns
- e) PRF: *0 to 5 kHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/200 V*
50-60 kHz

60 db ATTEN

100 V_{0.93} / DIV

TOP 10 ns / DIV

POI 1.0 ns / DIV
(RISE TIME)

PRF ≈ 5 kHz