


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMH-1-CP*

S.N.: *7850*

Date: *JAN 25 1997*



a) Output signal amplitude:  
*0 TO +10 VOLTS*

b) Pulse width:  
*≤ 200 PS*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF: *0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:  
*120/240V*  
*50-60A*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*PRF ≈ 10 MHz*