


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-6C-C-1-0P1-F1*


S.N.: *7879*

Date: *JAN 16 1997*




- a) Output signal amplitude:  
*0 TO +5 AMP*
- b) Pulse width:  
*50 NS TO 5-8 μS*  
*(10% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF: *0 TO 10 KHz*  
*(10% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*

Ⓐ MAINFRAMES out to 60db  
- 100 VOLT/DIV, 10 NS/DIV.



- g) Prime power: *120/240 V*  
*50-60 Hz.*

Ⓑ OUTPUT CURRENTS OF  
OUTPUT MODULE.  
*1-0 AMP/DIV.*



*MA*