


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-151B*
S.N.: *7881*
Date: *JAN 14 1997*



- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: *UR ≥ 50K*
NA
- c) Rise time:
≤ 145
- d) Fall time:
≤ 145
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
+ 24 VDC.
700 mA.
[Signature]

50V/DIV
5 ns/DIV
R_L = 50K.