


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2-C-BE03-R5-P-0P1-M*

S.N.: *7888*

Date: *FEB 28 1997*



- a) Output signal amplitude:
✓ 0 TO +2 AMP
- b) Pulse width:
✓ 17 NS
- c) Rise time:
✓ ≤ 2 NS
- d) Fall time:
✓ ≤ 2 NS
- e) PRF:
✓ 0 TO 20 KHz.
- f) Jitter, stability:
✓ OK

*MONITOR OUT TO 40dB
100V/DIV, 2AMP/DIV
5NS/DIV*

g) Prime power: *✓ 120/240V
50-60 Hz*

*AVO-2T-BE02B3 (CONNECTED)
VIA A 5 METER CABLE
1A459A IN SERIES WITH
1-Ω 4500 AS TEST
LOAD. PRF = 20 KHz*