


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV6-4A-P

S.N.: 7897

Date: JAN 30 1997



- a) Output signal amplitude:
+250 TO 650 VOLTS
- b) Pulse width:
 $\leq 5 \text{ NS}$
- c) Rise time:
 $\leq 1.5 \text{ NS}$
- d) Fall time:
 $\leq 3.0 \text{ NS}$
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
800 mA MAX

60 db ATTEN

100 VOLTS/DIV

PRF = 10 KHz