


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-AN-2-P-HTA*
S.N.: *7923*
Date: *MAR 17 1997*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
2 TO 50 NS
- c) Rise time:
- d) Fall time:
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
- g) Prime power: *OK*
+15 VOLTS DC

300 mA (MAX)
[Signature]

40 dB ATTN
10 VOLTS/DIV
TRR 5 NS/DIV
MRD 200 PS/DIV (RISE TIME)
BOT 200 PS/DIV (FALL TIME)