


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-C-P-URB*

S.N.: *7924*

Date: *FEB 10 1997*



- a) Output signal amplitude:  
*+ 50V (to 50V)*
- b) Pulse width:  
*20 to 100 ns*
- c) Rise time:  
*≤ 250 ps*
- d) Fall time: *(x 200 ps)*  
*≤ 2 ns*
- e) PRF:  
*0 to 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 db ATTEN*  
*'TOP' 5 ns/div*  
*≈ 32 VOLTS/DIV*

*120/240V*  
*50-60 Hz.*

*'BOT' 200 ps/div*  
*= 32 VOLTS/DIV*  
*(RISE TIME)*

*PRF = 20 kHz*

*[Signature]*